



623

Digital Manometer Instruction Manual



Table of Contents

Page 3	Introduction & Product Description
Page 4	Safety Considerations
Page 5	Technical Data & International Symbols
Page 6	Features & Benefits
Page 7	Pressure Conversion Table
Page 8	Step by Step Procedures
Page 10	Accessories & Specifications
Page 11	Maintenance
Page 12	Trouble shooting guide
Page 13	EC declaration of conformity

Introduction

Congratulations

Thank you for purchasing TPI brand products. This meter is easy to use and is built to last. It is backed by a 2 year limited warranty. Please remember to return your product warranty registration card.

Product Description

The TPI 623 is a hand held pressure meter with superior resolution .

Auto Range	-100.0 to -100.0 mbar -40.0 to 40.0 inH ₂ O
Resolution	0.001 inH ₂ O / 0.01 inH ₂ O
Accuracy	< +/-2 inH ₂ O: ±(0.02 inH ₂ O) > +/- 2 inH ₂ O: ±(1% of reading)
Over Pressure	Protected to 50 inH ₂ O
Power Supply	9V



Safety Considerations

WARNING: Please follow the manufacturer's test procedure whenever possible. Make sure the pressure to be measured is not outside the range of the TPI 623 meter.

ALWAYS

- Test the pressure meter before use to ensure it is operating properly
- Ensure that the connecting hose you are using is free from kinks or splits
- Double check all connections before testing

NEVER

- Connect the Instrument to an unknown pressure source if the source is twice the working pressure of the instrument the sensor will be damaged and may cause injury to the user
- Tamper with the instrument or attempt to modify it in any way. Otherwise, accident and injury may occur and the warranty becomes void.




Technical Data

Controls




Turns the 623 on. Hold down to turn unit off.

ZERO Momentarily press the  button to “ZERO” the display (“ZERO” will temporarily display in the lower left of the LCD)



Smoothing function to stabilize rapidly changing pressure readings

UNITS Selects 7 units of pressure kPa, PSI, inHg, mmHg, inH₂O, mmH₂O, mbar



Turns the back light on (automatically turns off after 30 seconds)



Features & Benefits

- Autoranging - Provides high resolution & accuracy at low
- Auto Power Off - Powers off after 40 minutes to conserve battery life.
- Trim Mode - Smoothing function to stabilize erratic readings. It displays the average value of 4 values previously measured.
- Zero Function - Provides immediate push button zeroing for Speed and accuracy.
- Units - Change the display value in any one of the following engineering units. kPa, PSI, inHg, mmHg, inH2O, mmH2O and mbar.
- Back Light - Provides easy reading of LCD in low light areas.



Pressure Coverstion Table

	Atmos	bar	inHg	inH ₂ O	PSI	mmHg	mmH ₂ O	Pascals
Atmos	1	9.86923 $\times 10^{-1}$	3.34207 $\times 10^{-2}$	2.458 $\times 10^{-3}$	6.8046 $\times 10^{-2}$	1.316 $\times 10^{-3}$	0.9677 $\times 10^{-4}$	9.869 $\times 10^{-6}$
bar	1.01325	1	3.3864 $\times 10^{-2}$	2.491 $\times 10^{-3}$	6.8948 $\times 10^{-2}$	1.333 $\times 10^{-3}$	0.9807 $\times 10^{-4}$	1.000 $\times 10^{-5}$
inHg	29.9213	29.53	1	7.355 $\times 10^{-2}$	2.036	3.937 $\times 10^{-2}$	2.89567 $\times 10^{-3}$	2.953 $\times 10^{-4}$
inH ₂ O	406.8	401.48	13.60	1	27.68	0.5354	3.937 $\times 10^{-2}$	4.014 $\times 10^{-3}$
PSI	14.6955	14.504	0.4912	3.6127 $\times 10^{-2}$	1	1.934 $\times 10^{-2}$	1.422323 $\times 10^{-3}$	1.4503 $\times 10^{-3}$
mmHg	760	750.06	25.400	1.868	51.715	1	7.35433 $\times 10^{-2}$	7.502 $\times 10^{-3}$
mmH ₂ O	10.3327 $\times 10^3$	10.1976 $\times 10^3$	3.4544 $\times 10^2$	25.400	703.07	13.599	1	.101972
Pascals	1.01325 $\times 10^5$	1.000 $\times 10^5$	3.386 $\times 10^3$	2.491 $\times 10^2$	6.8948 $\times 10^3$	1.333 $\times 10^2$	9.8071	1



Step by Step Procedures



Note:

The 623 incorporates a differential manometer. This means when a single hose is connected to the (+) port and pressure is applied, the 623 will read positive pressure. If a single hose is connected to the (-) port and pressure is applied, the 623 will read a negative pressure. If both ports are connected to a pressure, the 623 will read the difference ($P1 - P2$) between the two.

When performing single pressure measurements (non-differential) use the positive port and leave the negative port open.



Step by Step Procedures (continued)

1. Press the 'ON' key and the LCD display will be activated and automatically zeroed (if it does not read 0.00 proceed as follows)
2. Momentarily Press the  key and the display should zero within +/- 0.01 if not repeat until it does
3. Connect pressure meter to installation or equipment to be tested using appropriate hose connection in accordance with the manufacturers instructions
4. The pressure reading will appear on the display. To activate the backlight in poor light conditions press the  key at any time
5. If the reading is erratic or changing activate the smoothing function by pressing the '=' key
6. To convert the displayed pressure reading to another engineering unit depress the 'UNIT' key anytime and scroll through to the desired units of pressure.



Accessories & Specifications

Standard accessories

9V battery	A009
Soft Pouch	A600
Protective boot with magnets	A6XX
Silicone tubing w/ 1/4" barbed 1/8" NPT fitting.	A603

Specifications

Auto Range	-100.0 to -100.0 mbar -40.0 to 40.0 inH ₂ O
Resolution	0.001 inH ₂ O / 0.01 inH ₂ O
Accuracy	< +/-2 inH ₂ O: ±(0.02 inH ₂ O) > +/- 2 inH ₂ O: ±(1% of reading)
Over Pressure	Protected to 50 inH ₂ O
Power Supply	9V
Relative Humidity	0% to 80%
Operating Temp.	0°C to +40°C
Storage Temp.	-10°C to +50°C
Auto Power Off	20 Minutes
Input Type	Differential



Maintenance

Battery Replacement:

The 623 will display a battery symbol when the internal 9V battery needs replacement.

Replace battery as follows:

1. Ensure instrument is switched off
2. Loosen the six cross head screws on the back of the instrument
3. Remove the back cover
4. Remove old battery
5. Replace with new battery 9V NEDA 1604 or equivalent
6. Reassemble the instrument in reverse order from above being sure not to trap any wires between the two halves of the case
7. Re-fit the six cross head screws being sure not to over tighten

Cleaning: Use a mild detergent and a slightly damp cloth to clean the surfaces of the instrument.



Trouble Shooting Guide

Problem

Probable Cause

Unit does not zero

Zero button has not been pushed.

Unit still does not zero

If reading is greater than 0.05 and constantly fluctuates return to local Service Centre.

'Bat' appears on display

Replace battery immediately 9V NEDA 1505 or equivalent



EC Declaration of Conformity

This is to certify that model 623 conforms to the protection requirements of the council directive 89/336/EEC, in the approximation of laws of the member states relating to Electromagnetic compatibility by application of the following standards
EN50081-1 1992 Emissions Standard
EN50082-1 1992 Immunity Standard

To ensure conformity with these standards, this instrument must be operated in accordance with the instructions and specifications given in this manual.





Test Products International, Inc.

9615 SW Allen Blvd., Ste. 104

Beaverton, OR USA 97005

503-520-9197 • Fax: 503-520-1225

info@tpi-thevalueleader.com

L621M • 11/25/06 copyright © 2006 Test Products International, Inc.